

Product	Medical power supply for building-in
Applicant	EOS Power India PVT Ltd. Unit #57, SDF II, Seepz-Sez Andheri (East), Mumbai 400 096 India
Manufacturer	EOS Power India PVT Ltd. Unit #57, SDF II, Seepz-Sez Andheri (East), Mumbai 400 096 India
Factory	EOS Power India PVT Ltd. Unit #57, SDF II, Seepz-Sez Andheri (East), Mumbai 400 096 India <input type="checkbox"/> See next page(s)
Ratings	AC Input: 3.0 A Max 100-240 V~ 47-63Hz
Trade mark	
Model / Type Ref.	LFMWLP120-1XXX-YYY-W
Principal characteristics	Cl. I. Output Rating see the test report. In model name, the: "X" (first) = 0 or 3; "XX" (second and third) = 01 to 99; "Y" = 0-9, A-Z, a-z or blank, W is blank for MOPP or 1 for MOOP for Cl. I <input type="checkbox"/> See next page(s)
A sample of the product was tested and found to be in conformity with	MED EN 60601-1:2006;A11;A1;A12
Validity	This certificate documents conformity with the standards shown, and also applies as license for use of Nemkos name and certification mark. The certificate and license is valid as long as the applicable conditions are complied with, and provided that any changes to the product are notified to Nemko for acceptance prior to implementation. New standards or amendments to the standards may imply that the product design must be updated and/or that re-testing and re-certification is necessary.
Additional information	<input type="checkbox"/> See next page(s) The abovementioned certified equipment complies with current regulatory requirements regarding electrical safety in Norway and other EU/EEA member states, as far as this can be checked. Compliance with requirements regarding building-in, protection against electric shock and Electromagnetic Compatibility (EMC) must be checked when the equipment is built-in a completed product or forms a part of a complete system.
Additional model(s)	<input checked="" type="checkbox"/> See next page(s)

Date of issue 27-10-2015

Juan Z. Kleppenes
Certification Department**Nemko AS**Gaustadalléen 30, P.O. Box 73 Blindern, 0314 Oslo, Norway
TEL +47 22 96 03 30 FAX +47 22 96 05 50 EMAIL info@nemko.com
ENTERPRISE NUMBER NO974404532



Order No. 294295

CERTIFICATE

No. P15220391



Page 2/2

Product

Medical power supply for building-in

Pos. No

1

Model / Type Ref.

LFMWLP75-1XXX-YYY-W, LFMWLP65-1XXX-YYY-W or
LFMWLP45-1XXX-YYY-W

Trade mark (if different from page 1)

Rating

AC Input: 2.0 A Max, 100-240V~ 47-63Hz

Principal characteristics

Cl. I. Output Rating see the test report. In model name, the: "X" (first) = 0 or 3; "XX" (second and third) = 01 to 99; "Y" = 0-9, A-Z, a-z or blank, W is blank for MOPP or 1 for MOOP for Cl. I