

Product Medical Power Supply (for building-in)  
Applicant EOS Power India PVT Ltd.  
Unit #57, SDF II, Seepz-Sez, Andheri (E)  
Mumbai 400096  
India  
Manufacturer EOS Power India PVT Ltd.  
Unit #57, SDF II, Seepz-Sez, Andheri (E)  
Mumbai 400096  
India  
Factory EOS Power India PVT Ltd.  
Unit #57, SDF II, Seepz-Sez, Andheri (E)  
Mumbai 400096  
India  
 See next page(s)

Ratings 1.4A max. 100-120/200-240V AC 47-63Hz

Trade mark 

Model / Type Ref. MVLT60-ZXXXYYY

Principal characteristics Cl. I, Output: Refer to Additional information.  
Max. output power is 60W.  
The Z in the model designation can be any number 1 to 3. The XXX can be any number. The YYY can be any alphanumeric character or blank. Models may also be provided with an optional prefix of "LF", designate RoHS compliance.  
 See next page(s)

A sample of the product was tested and found to be in conformity with MED EN 60601-1:2006

Validity This certificate documents conformity with the standards shown, and also applies as license for use of Nemkos name and certification mark. The certificate and license is valid as long as the applicable conditions are complied with, and provided that any changes to the product are notified to Nemko for acceptance prior to implementation. New standards or amendments to the standards may imply that the product design must be updated and/or that re-testing and re-certification is necessary.

Additional information  Additional information on page 2

The abovementioned certified equipment complies with current regulatory requirements regarding electrical safety in Norway and other EU/EEA member states, as far as this can be checked. Compliance with requirements regarding building-in, protection against electric shock and Electromagnetic Compatibility (EMC) must be checked when the equipment is built-in a completed product or forms a part of a complete system.

Additional model(s)  See next page(s)

Date of issue 18-11-2011



Signature  
Hanne Yndestad  
Certification Department

Product Medical Power Supply (for building-in)  
Pos. No 1  
Model / Type Ref. MVLT40-ZXXXYYY  
Trade mark (if different from page 1)  
Rating 1.1A max. 100-120/200-240V AC 47-63Hz  
Principal characteristics Cl. I, Output: Refer to Additional information.  
Max. output power is 40W.  
The Z in the model designation can be any number 1 to 3. The XXX can be any number. The YYY can be any alphanumeric character or blank. Models may also be provided with an optional prefix of "LF", designate RoHS compliance.

## Additional information

List of models with output rating:

MVLT40-1200 : V1: 8A 5.1V  
MVLT40-1200S1 : V1: 8A 5.1V  
MVLT40-1201 : V1: 3.5A 12V  
MVLT40-1202 : V1: 2.7A 15V  
MVLT40-1202S27 : V1: 2.7A 15V  
MVLT40-1203 : V1: 1.7A 24V  
MVLT40-1204 : V1: 0.83A 48V  
MVLT40-3200 : V1: 6A 5.2V  
MVLT40-3201 : V1: 6A 5.2V; V2: 1A 24V; V3: 0.5A -12.8V  
MVLT40-3202 : V1: 6A 5.2V; V2: 1.5A 14.6V; V3: 0.5A -14.8V  
MVLT40-3203 : V1: 6A 5.2V; V2: 3A 14.6V; V3: 0.5A -14.8V

MVLT60-1200 : V1: 10A 5V  
MVLT60-1201 : V1: 5A 12V  
MVLT60-1201S15 : V1: 5A 12V  
MVLT60-1201S65 : V1: 0.1A 11.97V; 5A 10.5V  
MVLT60-1201S8 : V1: 5A 12V  
MVLT60-1202 : V1: 4A 15V  
MVLT60-1203 : V1: 1.7A 24V  
MVLT60-1204 : V1: 1.25A 48V  
MVLT60-1206 : V1: 4.22A 14.2V  
MVLT60-1207 : V1: 5A 13.2V  
MVLT60-1208 : V1: 2.92A 20.5V  
MVLT60-1209 : V1: 3.3A 18V  
MVLT60-3200 : V1: 8A 5.2V; V2: 3A 12.5V; V3: 0.5A -12.8V  
MVLT60-3201 : V1: 8A 5.2V; V2: 1.5A 24V; V3: 0.5A -12.8V  
MVLT60-3202 : V1: 8A 5.2V; V2: 2.5A 14.6V; V3: 0.5A -14.8V

Date of issue 18-11-2011



Signature

Hanne Yndestad

Certification Department