

Product Medical Power Supply (for building-in)
Applicant EOS Power India PVT.Ltd.
Unit #57, SDF II, Seepz-Sez, Andheri (E)
Mumbai - 400096
India

Manufacturer EOS Power India PVT.Ltd.
Unit #57, SDF II, Seepz-Sez, Andheri (E)
Mumbai - 400096
India

Factory EOS Power India PVT.Ltd.
Unit #57, SDF II, Seepz-Sez, Andheri (E)
Mumbai - 400096
India
 See next page(s)

Ratings 2.2A 100-240V AC 47-63Hz

Trade mark



Model / Type Ref. LFMWLT100-1XXX-YYY-WVU

Principal characteristics Cl. I, DC-outputs, please refer to Additional information.
The X in the model designation can be any number. The Y can be any alphanumeric character or blank. The W may be 1 or blank. The U may be 1 or blank. The LF is a prefix used to denote RoHS compliance.
 See next page(s)

A sample of the product was tested and found to be in conformity with MED EN 60601-1:2006

Validity This certificate documents conformity with the standards shown, and also applies as license for use of Nemkos name and certification mark. The certificate and license is valid as long as the applicable conditions are complied with, and provided that any changes to the product are notified to Nemko for acceptance prior to implementation. New standards or amendments to the standards may imply that the product design must be updated and/or that re-testing and re-certification is necessary.

Additional information Additional information on page 2

The abovementioned certified equipment complies with current regulatory requirements regarding electrical safety in Norway and other EU/EEA member states, as far as this can be checked. Compliance with requirements regarding building-in, protection against electric shock and Electromagnetic Compatibility (EMC) must be checked when the equipment is built-in a completed product or forms a part of a complete system.

Additional model(s) See next page(s)

Date of issue 19-04-2012



Signature
Hanne Yndestad
Certification Department

Product Medical Power Supply (for building-in)
Pos. No 1
Model / Type Ref. LFMWLT150-1XXX-YYY-WVU
Trade mark (if different from page 1)
Rating 2.2A 100-240V AC 47-63Hz
Principal characteristics Cl. I, DC-outputs, please refer to Additional information.
The X in the model designation can be any number. The Y can be any alphanumeric character or blank. The W may be 1 or blank. The U may be 1 or blank. The LF is a prefix used to denote RoHS compliance.

Additional information DC-output rating for the models.
The rated Aux output for the models below is 12V @ 0.5A, however the Aux output can be up to 24V (limited to 6 watts).
The different current ratings are for forced air cooling / conventional airflow cooling.

LFMWLT150-1000-W; LFMWLT100-1000-W : 20/16A 5V, 100/80W
LFMWLT150-1001-W; LFMWLT100-1001-W : 12.5/8.33A 12V, 150/100W
LFMWLT150-1002-W; LFMWLT100-1002-W : 10/6.67A 15V, 150/100W
LFMWLT150-1003-W; LFMWLT100-1003-W : 6.25/4.17A 24V, 150/100W
LFMWLT150-1004-W; LFMWLT100-1004-W : 3.13/2.08A 48V, 150/100W
LFMWLT150-1300-W; LFMWLT100-1300-W : 16/13A 5V, 150/100W
LFMWLT150-1301-W; LFMWLT100-1301-W : 12.5/8.33A 12V, 150/100W
LFMWLT150-1302-W; LFMWLT100-1302-W : 10/6.67A 15V, 150/100W
LFMWLT150-1303-W; LFMWLT100-1303-W : 6.25/4.17A 24V, 150/100W
LFMWLT150-1304-W; LFMWLT100-1304-W : 3.13/2.08A 48V, 150/100W

Date of issue 19-04-2012



Signature

Hanne Yndestad

Certification Department

Nemko AS	Office address	Internet	Telephone	Fax
P.O. Box 73, Blindern	Gaustadalléen 30	www.nemko.com	+47 22 96 03 30	+ 47 22 96 05 50
N-0314 Oslo, Norway	Oslo		Enterprise number:	NO 974404532