

Product	Medical Power Supply
Applicant	EOS Power India PVT Ltd. Unit #57, SDF II, Seepz-Sez Andheri (E) Mumbai - 400096 India
Manufacturer	EOS Power India PVT Ltd. Unit #57, SDF II, Seepz-Sez Andheri (E) Mumbai - 400096 India
Factory	EOS Power India PVT Ltd. Unit #57, SDF II, Seepz-Sez Andheri (E) Mumbai - 400096 India <input type="checkbox"/> See next page(s)
Ratings	1.9A 100-240Vac 47-63Hz
Trade mark	
Model / Type Ref.	MVC65NRXXYYY-WV
Principal characteristics	Cl. I, DC-output: 4.58A 12V 55W; 2.00A 15V 30W; 3.67A 15V 55W; 3.42A 19 65W; 2.7A 24V 65W; 1.35A 48V 65W The "XX" in the model name can be 12, 15, 19, 24 or 48, denoting output voltage. "YYY" may be any alphanumeric character or blank. "W" may be 1 or blank and "V" may be 2 or blank, denoting use of capacitors. Additionally models may be provided with an optional LF prefix used to designate RoHS compliance. <input type="checkbox"/> See next page(s)
A sample of the product was tested and found to be in conformity with	MED EN 60601-1:2006
Validity	This certificate documents conformity with the standards shown, and also applies as license for use of Nemkos name and certification mark. The certificate and license is valid as long as the applicable conditions are complied with, and provided that any changes to the product are notified to Nemko for acceptance prior to implementation. New standards or amendments to the standards may imply that the product design must be updated and/or that re-testing and re-certification is necessary.
Additional information	<input type="checkbox"/> See next page(s) The abovementioned certified equipment complies with current regulatory requirements regarding electrical safety in Norway and other EU/EEA member states, as far as this can be checked. Compliance with requirements regarding building-in, protection against electric shock and Electromagnetic Compatibility (EMC) must be checked when the equipment is built-in a completed product or forms a part of a complete system. <input type="checkbox"/> See next page(s)
Additional model(s)	
Date of issue 19-11-2012	
	
Signature	
Hanne Yndestad	
Certification Department	