

# CERTIFICATE OF COMPLIANCE

**Certificate Number** 20140203-E173812  
**Report Reference** E173812-A21-UL  
**Issue Date** 2014-February-03

**Issued to:** EOS POWER INDIA PVT LTD  
UNIT 57 SDF-II  
SEEPZ ANDHERI (E)  
MUMBAI MH 400096 INDIA

**This is to certify that  
representative samples of**

COMPONENT - POWER SUPPLIES, MEDICAL AND  
DENTAL

See Addendum page

Have been investigated by UL in accordance with the  
Standard(s) indicated on this Certificate.



**Standard(s) for Safety:**

ANSI/AAMI ES60601-1 (2005 + C1:09 + A2:10)(Medical  
Electrical Equipment - Part 1: General Requirements for  
Basic Safety and Essential Performance)  
CAN/CSA-C22.2 No. 60601-1 (2008) (Medical Electrical  
Equipment - Part 1: General Requirements for Basic Safety  
and Essential Performance)


**Additional Information:**

See the UL Online Certifications Directory at  
[www.ul.com/database](http://www.ul.com/database) for additional information

Only those products bearing the UL Recognized Component Marks for the U.S. and Canada should be considered as being covered by UL's Recognition and Follow-Up Service and meeting the appropriate U.S. and Canadian requirements.

The UL Recognized Component Mark for the U.S. generally consists of the manufacturer's identification and catalog number, model number or other product designation as specified under "Marking" for the particular Recognition as published in the appropriate UL Directory. As a supplementary means of identifying products that have been produced under UL's Component Recognition Program, UL's Recognized Component Mark: , may be used in conjunction with the required Recognized Marks. The Recognized Component Mark is required when specified in the UL Directory preceding the recognitions or under "Markings" for the individual recognitions. The UL Recognized Component Mark for Canada consists of the UL Recognized Mark for Canada:  and the manufacturer's identification and catalog number, model number or other product designation as specified under "Marking" for the particular Recognition as published in the appropriate UL Directory.

Recognized components are incomplete in certain constructional features or restricted in performance capabilities and are intended for use as components of complete equipment submitted for investigation rather



William R. Carney, Director, North American Certification Programs

UL LLC

Any information and documentation involving UL Mark services are provided on behalf of UL LLC (UL) or any authorized licensee of UL. For questions, please contact a local UL Customer Service Representative at [www.ul.com/contactus](http://www.ul.com/contactus)



# CERTIFICATE OF COMPLIANCE

**Certificate Number** 20140203-E173812  
**Report Reference** E173812-A21-UL  
**Issue Date** 2014-February-03

than for direct separate installation in the field. The final acceptance of the component is dependent upon its installation and use in complete equipment submitted to UL LLC.

Look for the UL Recognized Component Mark on the product.

This is to certify that representative samples of the product as specified on this certificate were tested according to the current UL requirements.

Medical Power Supply (for building-in)

Model: LFMWLT450-ZXXX-YYY-V-U



William R. Carney, Director, North American Certification Programs

UL LLC

Any information and documentation involving UL Mark services are provided on behalf of UL LLC (UL) or any authorized licensee of UL. For questions, please contact a local UL Customer Service Representative at [www.ul.com/contactus](http://www.ul.com/contactus)

