



**NOTICE OF AUTHORIZATION TO APPLY THE UL MARK**

May 23, 2007

Mr. Nigel D'Souza  
Celetronix Usa Inc  
2125-B Madera Rd  
Simi Valley, CA 93065  
United States

E-mail: [nigel\\_dsouza@jabil.com](mailto:nigel_dsouza@jabil.com)

Reference: File E173812      Project 07CA23313      P.O. Number Nigel Dsouza  
Product:    USR/CNR - AMEND MODEL DESIGNATION FOR MODELS MVLT60-ZXXX, MVLT40-ZXXX  
                 AND MVLT25 -ZXXX TO ADD OPTIONAL "YYY" SUFFIX AND OPTIONAL "LF" PREFIX

Dear Mr. D'Souza,

Any information and documentation involving UL Mark services are provided on behalf of Underwriters Laboratories Inc. (UL) or any authorized licensee of UL.

UL's investigation of your product has been completed under the above project number and the subject product was determined to comply with the applicable requirements.

This letter temporarily supplements the UL Follow-Up Services Procedure and serves as authorization to apply the UL Recognized Marking and/or Recognized Component Mark only at the factory under UL's Follow-Up Service Program to the subject product, which is constructed as described below:

Identical to the subject model, which was submitted to UL for this investigation. The UL Records covering the product will be in the Follow-Up Services Procedure, File E173812, Volume 1.

To provide the manufacturer with the intended authorization to use the UL Mark, the addressee must send a copy of this Notice and all attached material to each manufacturing location as currently authorized in File E173812, Volume 1.

This authorization is effective from the date of this Notice and only for products at the indicated manufacturing locations. Records in the Follow-Up Services Procedure covering the product are now being prepared and will be sent to the indicated manufacturing locations in the near future. Please note that Follow-Up Services Procedures are sent to the manufacturers only unless the Applicant specifically requests this document.

Products that bear the UL Mark shall be identical to those that were evaluated by UL and found to comply with UL's requirements. If changes in construction are discovered, appropriate action will be taken for products not in conformance with UL's requirements and continued use of the UL Mark may be withdrawn.

Sincerely,

Melissa J. DeGuia  
Project Engineer  
Department: 3011ACAM  
Tel: 360-817-5638  
Fax: 360-817-6162  
E-mail: [melissa.deguia@us.ul.com](mailto:melissa.deguia@us.ul.com)

Reviewed by:

Michael J. Howell  
Senior Project Engineer  
Department: 3016LSCL  
E-mail: [michael.j.howell@us.ul.com](mailto:michael.j.howell@us.ul.com)



## DESCRIPTION

## PRODUCT COVERED:

\*      USR, CNR: Component - Medical Power Supplies, Models MVLT60-ZXXXXYYY, MVLT40-ZXXXXYYY, MVLT25-ZXXXXYYY, where Z is any number 1 to 3, X is any Number, Y is any alphanumeric character or blank. Model designation may be provided with an optional prefix of "LF" designating RoHS compliance.

## ELECTRICAL RATINGS:

Input			Pins	Output		
V	A	Hz		V	A	W (max)
100-120/ 200-240	1.4	47-63	P1-2	1.5 to 3.3	11	33
			P1-2	3.3 to 12	10	50
			P1-2	12 to 58	5	60
			P5	-12 to -58	0.5	6
			P6	12 to 58	3	36

Maximum Output Power:   MVLT60-ZXXXXYYY = 60 W  
                                   MVLT40-ZXXXXYYY = 40 W  
                                   MVLT25-ZXXXXYYY = 25 W

## ENGINEERING CONSIDERATIONS (NOT FOR FIELD REPRESENTATIVE'S USE):

USR - indicates investigation to the Standard Medical Electrical Equipment, UL 60601-1, First Edition.

CNR - indicates investigation to CSA C22.2 No. 601.1

The products were tested on a 20 A branch circuit. If used on a branch circuit greater than this, additional testing may be necessary.