

Product Medical Power Supply (for building-in)
Applicant EOS Power India PVT.Ltd.
Unit #57, SDF II, Seepz-Sez, Andheri (E)
Mumbai - 400096
India

Manufacturer EOS Power India PVT.Ltd.
Unit #57, SDF II, Seepz-Sez, Andheri (E)
Mumbai - 400096
India

Factory EOS Power India PVT.Ltd.
Unit #57, SDF II, Seepz-Sez, Andheri (E)
Mumbai - 400096
India
 See next page(s)

Ratings 1.5A 100-240V AC 47-63Hz

Trade mark



Model / Type Ref. LFMWLT60-ZXXX-YYY-WV

Principal characteristics Cl. I, DC-outputs: Refer to Additional information page 2
The "Z" in model designation may be 1, 2 or 3, representing number of outputs. The "X" may be any numeric character, denoting configuration. The "Y" may be any alphanumeric character or blank, denoting minor output variation and/or minor SELV circuit variation. "W" may be 1 or blank and "V" may be 2 or blank.
 See next page(s)

A sample of the product was tested and found to be in conformity with MED EN 60601-1:2006

Validity This certificate documents conformity with the standards shown, and also applies as license for use of Nemkos name and certification mark. The certificate and license is valid as long as the applicable conditions are complied with, and provided that any changes to the product are notified to Nemko for acceptance prior to implementation. New standards or amendments to the standards may imply that the product design must be updated and/or that re-testing and re-certification is necessary.

Additional information Additional information on page 2

The abovementioned certified equipment complies with current regulatory requirements regarding electrical safety in Norway and other EU/EEA member states, as far as this can be checked. Compliance with requirements regarding building-in, protection against electric shock and Electromagnetic Compatibility (EMC) must be checked when the equipment is built-in a completed product or forms a part of a complete system.

Additional model(s) See next page(s)

Date of issue 17-01-2013

Hanne Yndestad

Signature

Hanne Yndestad

Certification Department

Nemko AS

Office address

Internet

Telephone

Fax

P.O. Box 73, Blindern

Gaustadalléen 30

www.nemko.com

+47 22 96 03 30

+ 47 22 96 05 50

N-0314 Oslo, Norway

Oslo

Enterprise number:

NO 974404532

Product Medical Power Supply (for building-in)
Pos. No 1
Model / Type Ref. LFMWLT40-ZXXX-YYY-WV
Trade mark (if different from page 1)
Rating 1.5A 100-240V AC 47-63Hz
Principal characteristics Cl. I, DC-outputs: Refer to list below.
The "Z" in model designation may be 1, 2 or 3, representing number of outputs. The "X" may be any numeric character, denoting configuration. The "Y" may be any alphanumeric character or blank, denoting minor output variation and/or minor SELV circuit variation. "W" may be 1 or blank and "V" may be 2 or blank.

Additional information

DC-outputs:

LFMWLT60-1000 : 10A 5.2V
LFMWLT60-1001 : 5.42A 12V
LFMWLT60-1002 : 4.33A 15V
LFMWLT60-1003 : 2.71A 24V
LFMWLT60-1004 : 1.35A 48V
LFMWLT60-1005 : 1.8A 36V

LFMWLT60-3000 : 8.0A 5.2V, 3.0A 12.5V, 0.5A -12.5V
LFMWLT60-3001 : 8.0A 5.2V, 1.5A 23.8V, 0.5A -12.5V
LFMWLT60-3002 : 8.0A 5.2V, 2.5A 14.6V, 0.5A -16.2V
LFMWLT60-3003 : 8.0A 3.3V, 3.0A 5.2V, 0.5A -12.8V

LFMWLT40-1000 : 8.0A 5.1V
LFMWLT40-1001 : 3.5A 12V
LFMWLT40-1002 : 2.7A 15V
LFMWLT40-1003 : 1.7A 24V
LFMWLT40-1004 : 0.83A 48V

LFMWLT40-3000 : 6.0A 5.2V, 2.0A 12.5V, 0.5A -12.8V
LFMWLT40-3001 : 6.0A 5.2V, 1.0A 23.8V, 0.5A -12.8V
LFMWLT40-3002 : 6.0A 5.2V, 1.5A 14.6V, 0.5A -14.8V
LFMWLT40-3003 : 6.0A 3.3V, 3.0A 5.2V, 0.5A -12.8V

This certificate replaces certificate issued 08-01-2013, due to misprint of model names in the DC-output list.

Date of issue 17-01-2013



Signature

Hanne Yndestad

Certification Department

Nemko AS

Office address

Internet

Telephone

Fax

P.O. Box 73, Blindern

Gaustadalléen 30

www.nemko.com

+47 22 96 03 30

+ 47 22 96 05 50

N-0314 Oslo, Norway

Oslo

Enterprise number:

NO 974404532