

Product	Medical Power Supply (for building-in)
Pos. No	5
Model / Type Ref.	MVPS600-1ZXX-YYYY-R
Trade mark (if different from page 1)	
Rating	Input: 6.5 A Max, 100-240 Vac, 47-63 Hz, Cl. I, Output: 600W max (Refer to General product information in report 360643 for output rating conditions and variable definition)
Principal characteristics	Where: Z = can be 'S', 'T' or 0 (Representing type of top cover, S= Slotted, T= Plain and 0= no top cover) X = any numeric from 0-9 (Representing the output voltage which can be any voltage in the range 12V to 58V, max current as stated in ref table) Y = any alpha, numeric or alphanumeric character or blank (denotes minor output variation and /or minor SELV circuit variations) R= with built-in Redundant diode (Optional)

Date of issue 29-01-2019



Okhyun Jeon
Certification Department

Nemko AS
Gautstadalleen 30, P.O. Box 73 Blindern, 0314 Oslo, Norway
TEL +47 22 96 03 30 FAX +47 22 96 05 50 EMAIL info@nemko.com
ENTERPRISE NUMBER