

CB TEST CERTIFICATE CERTIFICAT D'ESSAI OCProduct
Produit

Medical Power Supply (for building-in)

Name and address of the applicant
Nom et adresse du demandeurEOS Power India PVT Ltd.
Unit #57, SDF II, Seepz-Sez, Andheri (East)
Mumbai 400 096
IndiaName and address of the manufacturer
Nom et adresse du fabricantEOS Power India PVT Ltd.
Unit #57, SDF II, Seepz-Sez, Andheri (East)
Mumbai 400 096
IndiaName and address of the factory
Nom et adresse de l'usineEOS Power India PVT Ltd.
Unit #57, SDF II, Seepz-Sez, Andheri (East)
Mumbai 400 096
IndiaNote: When more than one factory, please report on page 2
Note: Lorsque il y plus d'une usine, veuillez utiliser la deuxième page Additional information on page 2Ratings and principal characteristics
Valeurs nominales et caractéristiques principales

4.2A 100-240V AC 47-63Hz, Cl. I

Trademark (if any)
Marque de fabrique (si elle existe)Type of Manufacturer's Testing Laboratories used
Type de programme du laboratoire d'essais constructeurModel / Type Ref.
Ref. De type

LFMWLT300-1XXX-YYY-WU

Additional information (if necessary may also be reported on page 2)
Les informations complémentaires (si nécessaire, peuvent être indiqués sur la deuxième page Additional information on page 2A sample of the product was tested and found to be in conformity with
Un échantillon de ce produit a été essayé et a été considéré conforme à la

IEC 60601-1(ed.3);am1

As shown in the Test Report Ref. No. which forms part of this Certificate
Comme indiqué dans le Rapport desais numéro de référence qui constitue partie de ce Certificat

294637

This CB Test Certificate is issued by the National Certification Body
Ce Certificat desai OC est établi par l'Organisme **National de Certification**Gaustadalléen 30
NO-0373 Oslo, Norway

Date: 21-12-2015

Signature: Juan Z. Kleppenes
Certification Department

Additional information(if necessary)**Information complémentaire (si nécessaire)**

Output: 325W max. In model name, X is any number denoting output configurations (output voltage and current ratings); Y may be any alphanumeric character or blank used to identify variants/derivatives of base model with no changes or with minor secondary changes which do not affect the safety construction, critical components list or output power rating; W may be 4 or blank: 4 denoting use of an Y2-type capacitors (C5 and C6) between primary and ground (1 MOOP), or 'blank' denoting use of Y1-type capacitors (C5 and C6) between primary and ground (1 MOPP). U may be 1 or 2 or 3: 1 denoting use of a Y1-type capacitor (C64) between secondary output and ground (1 MOPP), or 2 denoting use of Y2-type capacitor (C64) between secondary output and ground (1 MOOP), or 3 denoting the use of a standard decoupling capacitor (C59) or optional unpopulated capacitor on PCB.



Gaustadalléen 30
NO-0373 Oslo, Norway

Date: 21-12-2015

Signature: Juan Z. Kleppenes
Certification Department